



Unit A2  
The Bridge Business Centre  
Timothy's Bridge Road  
Stratford-upon-Avon,  
CV37 9HW  
England  
Tel: (01789) 293460  
(01789) 721577  
Fax: (01789) 262470  
VAT Reg. No. 551 3055 73

### **FAA Compliance for B & D Electromedical Devices**

To Whom it may concern,

In line with the requirements of US Dept. of Transportation (DOT) final rule 73 FR 27614, May 13, 2008, as modified by Corrective Notice of 74 FR 11469, March 18, 2009, respiratory assistive devices may be used onboard aircraft, without further testing by the carrier, provided they have been tested for electromagnetic compatibility (EMC) in accordance with RTCA/DO-160, Section 21, Category M.

A sample of the following devices manufactured by B & D Electromedical has been tested and found to comply with RTCA/DO-160, Section 21, conducted emissions 150kHz – 152MHz and radiated emissions 100Mhz – 6GHz to category M limits.

Test results available from [support@nippyventilator.com](mailto:support@nippyventilator.com).

These devices are therefore considered to be FAA compliant.

Nippy 3+ ventilator  
Nippy Junior+ ventilator  
Nippy ST+ ventilator  
Nippy S+ ventilator

Signed

A handwritten signature in black ink, appearing to read 'N.M. Davis', is written in a cursive style.

B & D Electromedical

28 July 2010